



(A division of Gradko International Ltd.) St. Martins House, 77 Wales Street Winchester, Hampshire SO23 0RH tel.: 01962 860331 fax: 01962 841339 e-mail:diffusion@gradko.co.uk

LABORATORY ANALYSIS REPORT NITROGEN DIOXIDE IN DIFFUSION TUBES BY U.V.SPECTROPHOTOMETRY

REPORT NUMBER L03269R BOOKING IN REFERENCE L03269 **DESPATCH NOTE** 34321 CUSTOMER Braunton Parish Council Attn: Tracey Lovell **Chaloners Road** Braunton EX33 2ES

DATE SAMPLES RECEIVED 04/05/2017

	Exposure Data						TOTAL	
Location	Sample Number	Date On	Date Off	Time (hr.)	μg/m ³ *	ppb *	µg NO₂	
London Inn Caen Street, Braunton	848774	31/03/2017	28/04/2017	673.50	36.02	18.80	1.76	
The Square Braunton	848775	31/03/2017	28/04/2017	673.50	45.91	23.96	2.25	

673.50

0.18

0.10

0.009

Laboratory Blank

Comment: Results are not blank subtracted Results have been corrected to a temperature of 293 K (20%)

	inperature of 235 it (20)			
Overall M.U.	±7.8%	Limit of Detection	0.010µgNO ₂	
Tube Preparation : 20% TEA / Water		Analysed on UV05 Camspec M550		
Analyst Name	Zoe Munday	Report Checked By	Adam Robinson	
Date of Analysis	16/05/2017	Date of Report	16/05/2017	

Analysis carried out in accordance with documented in-house Laboratory Method GLM7

The Diffusion Tubes have been tested within the scope of Gradko International Ltd. Laboratory Quality Procedures calculations and assessments involving the exposure procedures and periods provided by the client are not within the scope of our UKAS accreditation. Those results obtained using exposure data shall be indicated by an asterisk (*). Any queries concerning the data in this report should be directed to the Laboratory Manager Gradko International Ltd. This report is not to be reproduced, except in full, without the written permission of Gradko International Ltd. Page 1 of 1

Form LQF32b Issue 7 – Oct 2016

REPORT OFFICIALLY CHECKED

Report Number L03269R

0/10	1450 1 01 1	
	Gradko International Ltd	
This signate	ure confirms the authenticity of these results	
Signed	1 Cates	
	L. Gates, Laboratory Manager	